

**SRI LANKA STANDARD 1100 PART 4/ SECTION 1: 2018**  
**(ISO 15774: 2017)**  
**UDC 641**

**METHODS OF TEST FOR  
HEAVY METALS IN FOOD**  
**PART 4 – DETERMINATION OF HEAVY METALS IN  
ANIMAL AND VEGETABLE FATS AND OILS**  
**SECTION 1: Determination of cadmium content by direct  
graphite furnace atomic absorption spectrometry**

**SRI LANKA STANDARDS INSTITUTION**



**Sri Lanka Standard**  
**METHODS OF TEST FOR HEAVY METALS IN FOOD**  
**PART 4 – DETERMINATION OF HEAVY METALS IN ANIMAL AND**  
**VEGETABLE FATS AND OILS**  
**SECTION 1: Determination of cadmium content by direct graphite furnace atomic**  
**absorption spectrometry**

**SLS 1100 Part 4/ Section 1: 2018**  
**(ISO 15774: 2017)**

**Gr. C**

*Copyright Reserved*  
**SRI LANKA STANDARDS INSTITUTION**  
**17, Victoria Place**  
**Elvitigala Mawatha**  
**Colombo 08**  
**Sri Lanka**

Sri Lanka Standards are subject to periodical revision in order to accommodate the progress made by industry. Suggestions for improvement will be recorded and brought to the notice of the Committees to which the revisions are entrusted.

This Standard does not purport to include all the necessary provisions of a contract

© ISO 2017 - All right reserved.

© SLSI 2018

All right reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the SLSI.

**Sri Lanka Standard**  
**METHODS OF TEST FOR HEAVY METALS IN FOOD**  
**PART 4 – DETERMINATION OF HEAVY METALS IN ANIMAL AND**  
**VEGETABLE FATS AND OILS**  
**SECTION 1: Determination of cadmium content by direct graphite furnace atomic**  
**absorption spectrometry**

**NATIONAL FOREWORD**

This Sri Lanka Standard was approved by the Sectoral Committee on Food Products and was authorized for adoption and publication as a Sri Lanka Standard by the Council of the Sri Lanka Standards Institution on 2018-08-10

In order to accommodate a number of methods within the scope of one Standard, this Standard is published in several parts.

This part of the Standard focuses on the test methods of heavy metals in animal and vegetable fats and oils and consists of two sections. This section is identical with **ISO 15774: 2017** Animal and vegetable fats and oils – Determination of cadmium content by direct graphite furnace atomic absorption spectrometry, published by the International Organization for Standardization (ISO).

This section of the Standard describes a method for the determination of trace amounts of cadmium in all types of crude or refined edible oils and fats. Milk and milk products (or fat coming from milk and milk products) are excluded from the scope of this part of the Standard.

**Terminology and conventions:**

The text of the International Standard has been accepted as suitable for publication, without deviation, as a Sri Lanka Standard. However, certain terminology and conventions are not identical with those used in Sri Lanka Standards. Attention is therefore drawn to the following:

- a) Wherever the words “International Standard” appear referring to this Standard, it should be interpreted as “Sri Lanka Standard”.
- b) The comma has been used throughout as a decimal marker. In Sri Lanka Standards it is the current practice to use the full point on the base line as the decimal marker.
- c) Wherever page numbers are quoted, they are ISO page numbers.

**CROSS REFERENCES**

**International Standard**

**Corresponding Sri Lanka Standard**

ISO 3696, Water for analytical laboratory use –  
Specification and test methods

No corresponding Sri Lanka Standard

---

---

**Animal and vegetable fats and oils —  
Determination of cadmium content  
by direct graphite furnace atomic  
absorption spectrometry**

*Corps gras d'origines animale et végétale — Détermination de la  
teneur en cadmium par spectrométrie d'absorption atomique à four  
graphite*





## **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org



# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>1</b>
<b>5 Reagents</b> .....	<b>1</b>
<b>6 Apparatus</b> .....	<b>2</b>
<b>7 Procedure</b> .....	<b>3</b>
7.1 General.....	3
7.2 Treatment of samples, vegetable oil and standards.....	3
7.3 Preparation of apparatus.....	3
7.4 Determination.....	4
7.4.1 Graphite tube blank.....	4
7.4.2 Vegetable oil.....	4
7.4.3 Calibration of apparatus.....	4
7.4.4 Oil (liquid) samples.....	4
7.4.5 Fat samples with melting point 40 °C and higher.....	4
7.4.6 Number of determinations.....	5
<b>8 Expression of results</b> .....	<b>5</b>
<b>9 Precision</b> .....	<b>5</b>
9.1 Interlaboratory test.....	5
9.2 Repeatability.....	5
9.3 Reproducibility.....	5
<b>10 Test report</b> .....	<b>5</b>
<b>Annex A (informative) Results of interlaboratory test</b> .....	<b>6</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*.

This second edition cancels and replaces the first edition (ISO 15774:2000), which has been technically revised to exclude its applicability for fat coming from milk and milk products.

# Animal and vegetable fats and oils — Determination of cadmium content by direct graphite furnace atomic absorption spectrometry

## 1 Scope

This document describes a method for the determination of trace amounts (micrograms per kilogram) of cadmium in all types of crude or refined edible oils and fats.

Milk and milk products (or fat coming from milk and milk products) are excluded from the scope of this document.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

## 4 Principle

The oil or fat is incinerated and atomized in a suitable graphite tube furnace with a platform connected to an atomic absorption spectrometer, previously calibrated using standard solutions of an organo-compound of cadmium. The metal content is determined from the observed absorption at a wavelength of 228,8 nm. Palladium is added as a matrix modifier in order to prevent loss of cadmium during the thermal pretreatment.

## 5 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified.

**5.1 Water**, of grade 1 according to ISO 3696.

**5.2 Cyclohexane**.

**5.3 Hydrochloric acid**.

**5.4 Palladium chloride**.

**5.5 Matrix modifier**, 0,1 % (mass/volume) palladium solution.

Dissolve 0,167 g of palladium chloride (5.4) in 50 ml water (5.1) in a 100 ml volumetric flask (6.4), adding 1 ml hydrochloric acid (5.3) and making up to volume with water.

**5.6 Vegetable oil**, refined.

Any liquid edible oil is suitable. It shall be stored in a metal-free polyethylene bottle. The cadmium content of the oil shall not be greater than 0,2 µg/kg.

**5.7 Organometallic cadmium standard**, e.g. Conostan, 5 000 mg/kg<sup>1)</sup>.

**5.8 Standard stock solution**, of concentration 10 mg/kg cadmium, prepared by diluting 200 mg of the organometallic standard (5.7) with 100 g of vegetable oil (5.6).

**5.9 Standard working solutions**

Prepare daily working solutions containing 2,5 µg/kg, 5,0 µg/kg and 10,0 µg/kg of cadmium by diluting 25 mg, 50 mg and 100 mg, respectively, of the stock solution (5.8) with 100 g of vegetable oil (5.6).

**5.10 Argon**, of 99,99 % minimum purity.

## 6 Apparatus

**6.1 Polyethylene or polypropylene bottles**, of capacities 20 ml and 50 ml, metal free, with caps.

The bottles are made metal free in the following way. Clean the bottles thoroughly with warm nitric acid (2 mol/l). Rinse with distilled water and dry the bottles in a dust-free drying oven at about 80 °C.

**6.2 Micropipettor**, to deliver 10 µl and 20 µl.

**6.3 Pipettor tips**.

**6.4 Volumetric flask**, of capacity 100 ml.

**6.5 Electric oven**, capable of being maintained at 60 °C ± 2 °C.

**6.6 Atomic absorption spectrometer**, equipped with “peak area” mode and “autocalibrate” mode, together with an appropriate electrode-less discharge lamp (or hollow cathode lamp) and deuterium background corrector (or Zeeman atomic absorption spectrometer).

**6.7 Graphite furnace atomizer**, placed in the atomic absorption spectrometer (6.6), equipped with a control unit for temperature programming.

**6.8 Graphite tube**, uncoated.

**6.9 Platform**, pyrolytic.

---

1) Conostan, available from Continental Oil Company, Ponca City, Oklahoma, USA, is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product. Equivalent products may be used if it can be shown that they lead to the same results.

**6.10 Autosampler** for graphite furnace atomizer (optional), with polyethylene sample cups.

To use an autosampler equipped with a heating device, regulated at  $60\text{ °C} \pm 10\text{ °C}$ , fill sample cups with vegetable oil (5.6), standard working solutions (5.9) and oil samples. Fat samples with a melting point of  $40\text{ °C}$  and higher shall be diluted 1:1 (by mass) with blank oil (5.6). The volume of sample injected shall be  $15\text{ }\mu\text{l}$ .

To use an autosampler with no heating device, dilute 1:1 (mass/volume) the vegetable oil (5.6), the standard working solutions (5.9) and oil or fat samples with an organic solvent [e.g. cyclohexane (5.2)] at ambient temperature. The volume of sample injected shall be  $20\text{ }\mu\text{l}$ .

The use of an organic solvent [e.g. cyclohexane (5.2)] as a rinse for the autosampler is required.

## 7 Procedure

### 7.1 General

Environmental dust may contribute to background contamination. It is recommended that the analyses be carried out in a dust-free environment.

### 7.2 Treatment of samples, vegetable oil and standards

**7.2.1** Place all samples, standard working solutions (5.9) and vegetable oil (5.6) in the oven (6.5) set at  $60\text{ °C}$ .

If an autosampler is used, dilute 1:1 (mass/volume) all solutions (vegetable oil, standard working solutions, samples) with an appropriate solvent [e.g. cyclohexane (5.2)]. Homogenize well before pouring the solution into the sample cups of the autosampler.

**7.2.2** Shake the samples vigorously if the metal content of a sample is known to be outside the given range (5.9). Dilute (by mass) with vegetable oil (5.6).

### 7.3 Preparation of apparatus

**7.3.1** Switch on the atomic absorption spectrometer and the background correction (deuterium or Zeeman).

**7.3.2** In accordance with the manufacturer's instructions supplied with the spectrometer, adjust the lamp current, slit, wavelength and amplification. The required wavelength is  $228,8\text{ nm}$ .

**7.3.3** Optimize the position of the graphite furnace atomizer (6.7) in the atomic absorption spectrometer (6.6) and set the required programme on the control unit of the furnace. Place the platform (6.9) in the graphite tube (6.8).

7.3.4 Pretreat before each injection the pipettor tip (6.3) by pipetting and then discarding 10 µl of cyclohexane (5.2).

**Table 1 — Example of a programme for graphite furnace atomizer**

Step	Temperature °C	Ramp time s	Hold time s	Internal gas flow ml/min
Injection of matrix modifier				
1	200	30	30	300
Injection of oil sample				
0 <sup>a</sup>	60	0	20	0
1	200	30	30	300
2	650	60	40	300
3	1 600	0	5	0
4	2 700	1	3	50
<sup>a</sup> Extra temperature programming step only for fat samples with melting points over 40 °C.				

## 7.4 Determination

### 7.4.1 Graphite tube blank

Record the absorption, if any, of the graphite tube (6.8) and autozero this absorption.

### 7.4.2 Vegetable oil

By means of a micropipettor (6.2) or an autosampler (6.10), inject 20 µl of the matrix modifier (5.5) into the graphite furnace (6.7) and initiate the modifier temperature programme (step 1; see Table 1). Inject 10 µl of the vegetable solution (5.6) into the graphite furnace, initiate the temperature programme (steps 1 to 4) and record the absorption (for the autosampler, see 6.10).

### 7.4.3 Calibration of apparatus

By means of a micropipettor (6.2), inject 20 µl of the matrix modifier (5.5) into the graphite furnace (6.7) and initiate the modifier temperature programme (step 1). Inject 10 µl of the first of the three standard working solutions, prepared according to 5.9, into the graphite furnace. Initiate the temperature programme (steps 1 to 4). Continue with the second and third standard working solutions successively. Calibrate the spectrometer according to the operating procedure of the apparatus used (for the autosampler, see 6.10). Plot the analytical curve.

### 7.4.4 Oil (liquid) samples

By means of a micropipettor (6.2), inject 20 µl of the matrix modifier (5.5) into the graphite furnace (6.7) and initiate the modifier temperature programme (step 1). Inject 10 µl of the oil sample into the graphite furnace, initiate the temperature programme (steps 1 to 4) and record the concentration according to the operating procedure of the apparatus used (for the autosampler, see 6.10).

### 7.4.5 Fat samples with melting point 40 °C and higher

Introduce an extra temperature-programming step with temperature 60 °C, hold time 20 s, and with an internal gas flow of 0 ml/min. By means of a micropipettor (6.2), inject 20 µl of the matrix modifier (5.5) into the graphite furnace (6.7) and initiate the modifier temperature programme. In the first programme step (step 0 in Table 1), inject 10 µl of the melted fat into the graphite furnace by allowing the tip to remain in the injection opening to liquefy the fat before injecting it. Record the concentration according to the operating procedure of the apparatus used (for autosampler, see 6.10).

#### 7.4.6 Number of determinations

Carry out two determinations in rapid succession.

### 8 Expression of results

The measured concentration is expressed in micrograms per kilogram. Report as the final result the mean of the results of two determinations.

## 9 Precision

### 9.1 Interlaboratory test

Details of an interlaboratory test on the precision of the method are summarized in [Annex A](#). The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given.

### 9.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than:

- 1,1 µg/kg when the mean value lies between 2 µg/kg and 8 µg/kg for oil samples;
- 1,7 µg/kg when the mean value lies between 3 µg/kg and 7 µg/kg for fat samples.

### 9.3 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than:

- 2,9 µg/kg when the values of the two results lie between 2 µg/kg and 8 µg/kg for oil samples;
- 2,6 µg/kg when the values of the two results lie between 3 µg/kg and 7 µg/kg for fat samples.

## 10 Test report

The test report shall specify the following:

- all information necessary for the complete identification of the sample;
- the sampling method used, if known;
- the test method used, with reference to this document, i.e. ISO 15774;
- all operating details not specified in this document, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- the test result(s) obtained, or, if the repeatability has been checked, the final quoted result obtained.

## Annex A (informative)

### Results of interlaboratory test

An international collaborative test, carried out in 1994 by the IUPAC Commission on Oils, Fats and Derivatives, in which eight laboratories participated, gave the statistical data shown in [Table A.1](#).

**Table A.1 — Precision data**

	Sunflower oil		Lard	
	A	B	A	B
Number of laboratories retained after eliminating outliers	6	7	6	7
Number of outliers (laboratories)	2	1	2	1
Mean value, µg/kg	7,87	1,58	7,17	2,70
True value, µg/kg	8,0	2,0	7,0	3,0
Repeatability standard deviation, $s_r$ , µg/kg	0,31	0,39	0,61	0,23
Repeatability coefficient of variation, %	3,9	24,9	8,6	8,5
Repeatability limit, $r$ ( $2,8 \times s_r$ ), µg/kg	0,87	1,11	1,74	0,65
Reproducibility standard deviation, $s_R$ , µg/kg	0,99	0,55	0,92	0,59
Reproducibility coefficient of variation, %	13,2	34,8	12,9	22,0
Reproducibility limit, $R$ ( $2,8 \times s_R$ ), µg/kg	2,94	1,55	2,61	1,68







## **SLS CERTIFICATION MARK**

*The Sri Lanka Standards Institution is the owner of the registered certification mark shown below. Beneath the mark, the number of the Sri Lanka Standard relevant to the product is indicated. This mark may be used only by those who have obtained permits under the SLS certification marks scheme. The presence of this mark on or in relation to a product conveys the assurance that they have been produced to comply with the requirements of the relevant Sri Lanka Standard under a well designed system of quality control inspection and testing operated by the manufacturer and supervised by the SLSI which includes surveillance inspection of the factory, testing of both factory and market samples.*

*Further particulars of the terms and conditions of the permit may be obtained from the Sri Lanka Standards Institution, 17, Victoria Place, Elvitigala Mawatha, Colombo 08.*



## **SRI LANKA STANDARDS INSTITUTION**

The Sri Lanka Standards Institution (SLSI) is the National Standards Organization of Sri Lanka established under the Sri Lanka Standards Institution Act No. 6 of 1984 which repealed and replaced the Bureau of Ceylon Standards Act No. 38 of 1964. The Institution functions under the Ministry of Science & Technology.

The principal objects of the Institution as set out in the Act are to prepare standards and promote their adoption, to provide facilities for examination and testing of products, to operate a Certification Marks Scheme, to certify the quality of products meant for local consumption or exports and to promote standardization and quality control by educational, consultancy and research activity.

The Institution is financed by Government grants, and by the income from the sale of its publications and other services offered for Industry and Business Sector. Financial and administrative control is vested in a Council appointed in accordance with the provisions of the Act.

The development and formulation of National Standards is carried out by Technical Experts and representatives of other interest groups, assisted by the permanent officers of the Institution. These Technical Committees are appointed under the purview of the Sectoral Committees which in turn are appointed by the Council. The Sectoral Committees give the final Technical approval for the Draft National Standards prior to the approval by the Council of the SLSI.

All members of the Technical and Sectoral Committees render their services in an honorary capacity. In this process the Institution endeavours to ensure adequate representation of all view points.

In the International field the Institution represents Sri Lanka in the International Organization for Standardization (ISO), and participates in such fields of standardization as are of special interest to Sri Lanka.