SRI LANKA STANDARD 1255 : PART 7 : 2003 ISO 7176 – 7 : 1998

METHODS OF TEST FOR NON - FOLDING WHEELCHAIRS PART 7: MEASUREMENT OF SEATING AND WHEEL DIMENSIONS

SRI LANKA STANDARDS INSTITUTION

METHODS OF TEST FOR NON - FOLDING WHEELCHAIRS PART 7: MEASUREMENT OF SEATING AND WHEEL DIMENSIONS [WHEELCHAIRS-MEASUREMENT OF SEATING AND WHEEL DIMENSIONS]

SLS 1255 : Part -7 : 2003 ISO 7176-7 : 1998

Gr. T

SRI LANKA STANDARDS INSTITUTION
No. 17, Victoria Place
Off Eelvitigala Mawahta
Colombo 08
Sri Lanak

SLS 1255 : Part 7 : 2003

ISO 7176-7:1998

NATIONAL FOREWORD

This Sri Lanka Standard was authorized for adoption and publication by the Council of the Sri Lanka Standards Institution on 2003-12-19 after it had been approved by the Sectoral Committee on Materials, Mechanical Systems and Manufacturing Engineering.

This Sri Lanka Standards is identical with ISO 7176–7:1999 "Wheelchairs - Measurement of seating and wheel dimensions" published by the International Organization for Standardization (ISO).

This standard is published in eight parts as follows: SLS 1255 Methods of test for non-folding wheelchairs

- Part 1 Determination of static stability
- Part 3 Determination of effectiveness of breaks
- Part 5 Determination of overall dimensions, mass and turning space
- Part 7 Measurement of seating and wheel dimensions
- Part 8 Requirements and test methods for static, impact and fatigue strengths
- Part 11 Test dummies
- Part 13 Determination of coefficient of friction of test surfaces
- Part 22 Set up procedures

Terminology and Convention

The text of this International Standards has been accepted as suitable for publication without deviation, as a Sri Lanka Standard. However, certain terminology and conventions are not identical with those used in Sri Lanka Standards, and hence the attention is drawn to the following:

- i) Wherever the words 'International Standard' appear, referring to this standards they should be read as 'Sri Lanka Standard'.
- ii) The comma has been used throughout as a decimal marker. In Sri Lanka Standards, the current practice is that a full point on the base line is used as the decimal marker.

In reporting the result of a test or an analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with CS 102.

INTERNATIONAL STANDARD

ISO 7176-7

First edition 1998-05-15

Wheelchairs —

Part 7:

Measurement of seating and wheel dimensions

Fauteuils roulants —

Partie 7: Mesurage des dimensions d'assise et des roues



ISO 7176-7:1998(E)

Contents		Page
1	Scope	1
2	Normative references	1
3	Definitions and abbreviations	2
4	Principle	6
5	Test equipment	6
6	Preparation of the wheelchair	6
7	Measurement procedure	8
8	Measurement records and disclosure	41
Anr	nex A: Reference loader gauges	42
Annay R: Data forms		47

© ISO 1998

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet central@iso.ch
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland